


LABORATORY  
MANUAL  
CHEMISTRY

دليل المختبرات  
والأجهزة العلمية



إعداد لجنة الأجهزة والمعامل - 1444 هـ

## المحتويات

الصفحة	الموضوع
2	مقدمة
3	كلمة سعادة رئيس القسم
4	نبذة عن قسم الكيمياء بكلية العلوم
5	رؤية ورسالة وأهداف القسم
6	السلامة في المختبرات
8	مختبرات شطر الطلاب
9	مختبرات شطر الطالبات
10	المقررات العملية بقسم الكيمياء
11	التجارب العملية بقسم الكيمياء - الكيمياء العامة
13	التجارب العملية بقسم الكيمياء - الكيمياء العضوية
28	التجارب العملية بقسم الكيمياء - الكيمياء التحليلية
37	التجارب العملية بقسم الكيمياء - الكيمياء الفيزيائية
43	التجارب العملية بقسم الكيمياء - الكيمياء غير العضوية
48	دليل الأجهزة المستخدمة بمختبرات القسم



## مقدمة

تم بحمد الله إعداد هذا الدليل مشتملاً على قاعدة بيانات مبسطة لمختبرات قسم الكيمياء ومختلف الأجهزة العلمية بها والتي تخدم كافة المقررات التعليمية والأنشطة البحثية بالقسم. ويأتي إعداد هذه القاعدة وجمع كل الأجهزة في كتيب واحد ليكون دليلاً للطلاب والباحث ومرشداً لكل العاملين بالمختبر ليساهم في تطوير وتيسير الجلسات العملية وأبحاث التخرج وهي أحد أهم محاور العملية التعليمية بالقسم. وختاماً تتقدم لجنة الأجهزة والمعامل بالقسم بخالص التقدير لكل من ساهم في إخراج هذا الدليل سائلين المولى عز وجل أن يكون نافعاً ومفيداً لأبنائنا الطلاب والطالبات والعاملين بمختبرات القسم.

## لجنة الأجهزة والمعامل بالقسم



## كلمة سعادة رئيس القسم



تشهد مملكتنا حالياً تنمية شاملة النهضة ، أحد جوانبها الرئيسية هو التعليم العالي ، والذي يشهد حالياً نقلة نوعية في جميع المحاور ولقد تم إنشاء قسم الكيمياء بكلية العلوم بجامعة العريقة كأحدى سمات هذه النهضة. يأتي قسم الكيمياء في طليعة أقسام الجامعة التي تهدف إلى تأهيل خريجين ذوي كفاءات عالية للمساهمة في تلك النهضة الفعالة وتلبية احتياجات سوق العمل. يقوم بالتدريس والبحث في القسم نخبة من أعضاء هيئة التدريس الأكفاء المؤهلين لتقديم خدمات تنافسية على المستويين الأكاديمي والبحثي والمساهمة في حل مشكلات قطاعي الصحة والصناعة بتقديم المشورة العلمية المطلوبة لحل المشكلات التي تواجههم. يقوم القسم أيضاً بتقديم خدمات للأقسام المناظرة وذلك بالإشراف الأكاديمي على بعض مقررات الكيمياء بتلك الأقسام مثل كلية الهندسة وبرنامج المسار الصحي بالجامعة كما يعمل القسم حالياً على تطوير البرامج والخطط الدراسية لدرجة البكالوريوس بالإضافة إلى التخطيط لتقديم برنامج الدراسات العليا.

**د. وليد بن محمد الأمير**

**رئيس قسم الكيمياء**

## قسم الكيمياء بكلية العلوم



تم انشاء قسم الكيمياء عام 1426هـ تلبية للإنطلاقة التي تخطوها الجامعة نحو الريادة العلمية والتقدم، لمواكبة الاحتياجات الوطنية للكوادر العلمية المتخصصة والمدرّبة في مجالات الكيمياء المتعددة. ويعتبر قسم الكيمياء من أكبر اقسام كلية العلوم وتمددت خدمات التدريس به إلى قطاع كبير من طلبة الجامعة في عدد من البرامج خارج كلية العلوم مثل كلية الهندسة وكليات القطاع الصحي بالإضافة إلى الأنشطة البحثية وعمل شركات متميزة مع العديد من الشركات و المصانع وأصحاب الأعمال على مستوى المملكة حيث تقدم لهم المشورة العلمية المطلوبة لحل المشكلات التي تواجههم. يستخدم القسم التقنيات الحديثة في التدريس والمبنيّة على التعليم الإلكتروني واستخدام التطبيقات الحديثة في تفعيل التواصل المستمر مع الطلاب و تقديم الدعم العلمي والإرشادي بما يتواءم مع رؤية وأهداف القسم.

## رؤية القسم:

- تحقيق التميز في التعليم والبحث العلمي وخدمة المجتمع

## رسالة القسم:

- تقديم خدمات تعليمية وبخثية وبناء شركات فاعلة لخدمة المجتمع في الكيمياء

## أهداف القسم:

- إتاحة تعليم شامل وفعال قائم على المرافق و الخدمات بالقسم.
- تعزيز الاستفادة من تقنية المعلومات بالقسم.
- تفعيل وانجاز الخطط السنوية للقسم .
- تطوير قدرات ومهارات الموارد البشرية بالقسم.
- اتاحة برنامج أكاديمي بجودة عالية قائمة علي استراتيجيات تعليمية حديثة
- زيادة دعم وتطوير البحث العلمي والابتكار في مجالات الكيمياء المتعددة.
- تفعيل وتعزيز الشراكات التعليمية و البحثية والتدريبية بالقسم.
- تطوير مخرجات تعلم البرنامج وتنمية مهارات اعضاء هيئة التدريس والطلاب
- تحسين كفاءة خريجي القسم.
- تطوير دور القسم في خدمة المجتمع



## مختبرات القسم



تنقسم مختبرات قسم الكيمياء إلى خمسة تخصصات وهي الكيمياء العامة، والكيمياء العضوية، والكيمياء التحليلية، والكيمياء الفيزيائية. والكيمياء غير العضوية و يستوعب المختبر حوالي 20 طالباً/طالبة. يتم تدريس التجارب العملية بالمختبرات والتي تنسم بالبساطة والمتعة. يوجد بجميع مختبرات القسم مواد كيميائية صلبة أو سائلة كما أن بها مصادر للغاز و المياه و الكهرباء. ونتيجة لقلّة الوعي وعدم الامتثال لمعايير الصحة والسلامة يتعرض البعض للإصابات والأمراض المهنية و نظراً لأن الانطباع الذي يأخذه الطالب عن المختبر سيبقى أثره كبيراً خلال حياته العملية بعد مرحلة الدراسة فإنه من الواجب العمل بكل اشتراطات السلامة أثناء العمل بالمختبر.

### السلامة في المختبر

يعد مختبر الكيمياء مكاناً للتجريب والتعلم. لذا على الشخص أن يتحمل مسؤولية سلامته الشخصية وسلامة من يعملون بالقرب منه. الحوادث عادةً يسببها الإهمال، إلا أنه يمكن المساعدة على منعها بالاتباع الدقيق للتعليمات المتضمنة في هذا دليل السلامة بالمختبرات، وفيما يلي بعض قواعد السلامة التي تساعد على حماية النفس والآخرين من التعرض للإصابات في المختبر:

1. لا تتواجد في المختبر لوحده.
2. ادرس التجربة جيداً قبل مجيئك إلى المختبر.

3. البس النظارة الواقية، وارتداء معطف المختبر في أي وقت تعمل فيه في المختبر. كما يجب ارتداء القفازين كل مرة تستعمل فيها المواد الكيميائية لأنها تسبب التهيج، وقد يمتصها الجلد.
4. يُحظر وضع عدسات لاصقة في المختبر، حتى لو كنت تلبس نظارة واقية، فالعدسات تمتص الأبخرة، ويصعب إزالتها في الحالات الطارئة..
5. تجنب لبس الملابس الفضفاضة، فالملابس الفضفاضة قد تشتعل، كما أنها تشتبك بالأدوات المخبرية، وكذلك الحلي.
6. البس أحذية مغلقة تغطي القدم تمامًا؛ فالأحذية المكشوفة غير مسموح بها في المختبر.
7. اعرف مكان طفاية الحريق، ورشاش الماء، ومغسلة العينين، وبطانية الحريق، وصيدلية الإسعاف الأولية، واعرف أيضًا كيف تستعمل أدوات السلامة المتوفرة.
8. تعامل مع المواد الكيميائية بحذر، وتفحص بطاقة المعلومات التي على العبوات قبل أخذ أي كميات منها؛ خصوصًا قبل استعمالها.
9. لا ترجع المواد الكيميائية الفائضة إلى عبواتها الأصلية.
10. لا تأخذ عبوات المواد الكيميائية إلى مكان عملك واستعمل أنابيب اختبار، أو أوراقًا، أو كؤوسًا للحصول على ما يلزمك منها.
11. خذ كميات قليلة فقط لأن الحصول على كمية إضافية أسهل من التخلص من الفائض.
12. لا تدخل القطارات في عبوات المواد الكيميائية مباشرة، بل اسكب قليلًا منها في كأس.
13. لا تتذوق أو تشم أي مادة كيميائية أبدًا.
14. يُمنع الأكل والشرب ومضع العلكة في المختبر.
15. استعمل مائة الماصة عند سحب المواد الكيميائية، ولا تسحبها بفمك أبدًا.
16. إذا لامست مادة كيميائية عينيك أو جلدك فاغسلها مباشرةً بكميات كبيرة من الماء، وأخير أحد المختصين المتواجدين معك في المختبر.
17. احفظ المواد القابلة للاشتعال بعيدًا عن اللهب (الكحول مواد سريعة الاشتعال)
18. لا تتعامل مع الغازات السامة والقابلة للاحتراق إلا تحت إشراف مختص. واستعمل مثل هذه المواد داخل خزانة الغازات.
19. عند تسخين مادة في أنبوب اختبار كن حذرًا، فلا تُوجّه فوهة الأنبوب تجاه جسمك أو تجاه شخص آخر، ولا تنظر أبدًا في فوهة الأنبوب.
20. توجّه الحذر، واستعمل أدوات مناسبة عند الإمساك بالزجاج والأجهزة الساخنة. الزجاج الساخن لا يختلف في مظهره في الزجاج البارد.
21. تعرّف الطريقة الصحيحة لتحضير محاليل الأحماض، وأضف دائمًا الحمض ببطء إلى الماء.

22. حافظ على كفة الميزان نظيفة، ولا تضع أبدًا المواد الكيميائية في كفة الميزان مباشرةً.
23. لا تسخن المخابير المدرجة، أو السحاحات، أو الماصات باستعمال اللهب.
24. بعد أن تكمل التجربة، نظّف الأدوات وأعدّها إلى أماكنها، ونظّف مكان العمل، وتأكد من إغلاق مصادر الغاز والماء، واغسل يديك بالماء والصابون قبل أن تغادر المختبر.
25. تعرف على مخارج الطوارئ بالمختبر للخروج الآمن في حالة الحوادث أو الحرائق لا قدر الله.

## تصنيف مختبرات القسم :

تشمل مختبرات القسم على عدد (21) مختبر موزعة على مختلف التخصصات منها عدد (10) مختبرات بشطر الطلاب وعدد (11) مختبر بشطر الطالبات كما هو موضح بالجدول التالي:

### أولاً: مختبرات شطر الطلاب

م	اسم المختبر	رقم المختبر	موقع المختبر	مشرف المختبر	Lab.
1	معمل الكيمياء العضوية	G201	النور الأرضي	د. محمد صبري	Organic Chemistry Lab
2	معمل الكيمياء التحليلية	G203	النور الأرضي	د. محمد عطيف	Analytical Chemistry Lab.
3	معمل الكيمياء غير العضوية	G207	النور الأرضي	د. علي سليمان	Inorganic Chemistry Lab.
4	معمل الكيمياء الفيزيائية	G210	النور الأرضي	أ.د. حسام سيد	Physical Chemistry Lab.
5	معمل الكيمياء العضوية (2)	G703	النور الأرضي	د. سامي قرني	Organic Chemistry Lab. (2)
6	معمل الكيمياء التحليلية (2)	G508	النور الأرضي	د. محمد عطيف	Analytical Chemistry Lab. (2)
7	معمل الكيمياء الفيزيائية (2)	1712	النور الأول	د. وليد الأمير	Physical Chemistry Lab. (2)
8	معمل الكيمياء العضوية (3)	1704	النور الأول	د. سامح الجوجري	Organic Chemistry Lab. (3)
9	معمل الكيمياء العامة (1)	G602	النور الأرضي	د. أسامة سليمان	General Chemistry Lab. (1)
10	معمل الكيمياء العامة (2)	G605	النور الأرضي	د. أسامة سليمان	General Chemistry Lab. (2)

## ثانياً: مختبرات شطر الطالبات

Lab.	مشرفة المختبر	موقع المختبر	رقم المختبر	اسم المختبر	م
Physical Chemistry Lab.	م / دعاء محمد شريف	مبنى المعامل (المر الثاني)	DS-02	معمل الكيمياء الفيزيائية	1
General Chemistry Lab. (1)	د / هالة إسماعيل الزبير	مبنى المعامل (المر الثاني)	DS-03	معمل الكيمياء العامة (1)	2
General Chemistry Lab. (2)	د / لينا صلاح ابراهيم	مبنى المعامل (المر الثاني)	DS-04	معمل الكيمياء العامة (2)	3
General Chemistry Lab. (3)	د / نورا الإياري	مبنى المعامل (المر الثاني)	DS-05	معمل الكيمياء العامة (3)	4
Inorganic Chemistry Lab.	د / اميرة عبد المقصود	مبنى المعامل (المر الثاني)	DS-06	معمل الكيمياء غير العضوية	5
Organic Chemistry Lab. (1)	د / افراح محمد شرف	مبنى المعامل (المر الثاني)	DS-07	معمل الكيمياء العضوية (1)	6
Organic Chemistry Lab. (2)	م / ايمان علي باز	مبنى المعامل (المر الثاني)	DS-08	معمل الكيمياء العضوية (2)	7
Analytical Chemistry Lab. (1)	م / امنة مهدي حمدي	مبنى المعامل (المر الثاني)	DS-10	معمل الكيمياء التحليلية (1)	8
Analytical Chemistry Lab. (2)	د / سلمية ابراهيم يعقوب	مبنى المعامل (المر الثاني)	DS-11	معمل الكيمياء التحليلية (2)	9
Biochemistry Lab.	م / سارة حسين الطيب	مبنى المعامل (المر الثاني)	DS-13	معمل الكيمياء الحيوية	10
Physical Chemistry Lab.	د / زيزي البدرابي	مبنى المعامل (المر الثاني)	DS-16	معمل الكيمياء الفيزيائية	11

## المقررات العملية بقسم الكيمياء

عدد الوحدات المعتمدة	عدد الساعات الدراسية		اسم المقرر	رقم ورمز المقرر	م
	(عملي)	(نظري)			
4	2	3	كيمياء عامة	101كيم	1
4	2	3	كيمياء عامة وفيزيائية	201 كيم	2
3	2	2	كيمياء التحليل الحجمي	211 كيم	3
3	2	2	كيمياء التحليل الوزني	212 كيم	4
4	2	3	كيمياء المجموعات الرئيسية	221 كيم	5
3	2	2	كيمياء عضوية اليقاتية	231 كيم	6
3	2	2	كيمياء عضوية أروماتية	232 كيم	7
3	2	2	ديناميكا حرارية	241كيم	8
3	2	2	التحليل الكروماتوجرافي	313كيم	9
3	2	2	طرق التحليل الكهربائي	314 كيم	10
4	2	3	كيمياء العناصر الانتقالية	322كيم	11
3	2	2	كيمياء تناسقية وعضو فلزية	323كيم	12
3	2	2	كيمياء عضوية حلقيه غير متجانسة	333كيم	13
3	2	2	ميكانيكية التفاعلات العضوية	335كيم	14
3	2	2	كيمياء حركية	342كيم	15
3	2	2	كيمياء كهربية	344كيم	16
4	2	3	طرق التحليل الآلي	415 كيم	17
3	2	2	كيمياء اللانثانيدات والأكتينيدات	424كيم	18
3	2	2	كيمياء النواتج الطبيعية	436كيم	19
3	2	2	كيمياء عضوية تطبيقية	438كيم	20
3	2	2	مبادئ الكيمياء الحيوية	439كيم	21
3	2	2	كيمياء المحاليل	445كيم	22
2	2	1	بحث تخرج	491كيم	23

## Experiments in Chemistry Labs

1. General Chemistry
2. Organic Chemistry
3. Physical Chemistry
4. Inorganic Chemistry
5. Analytical Chemistry



### 1. Practical General Chemistry

#### 1.1 General Chemistry (101CHEM-4)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Safety and regulations	Tubes, pipet, beakers, bottles,...etc
2.	Experiment no. 1 Group 1 acidic radical	Dil HCl, $\text{CO}_3^{2-}$ , $\text{HCO}_3^-$ , $\text{S}_2\text{O}_3^{2-}$ , $\text{BaCl}_2$ , $\text{MgSO}_4$
3.	Experiment no. 2 Group 2 acidic radical	Conc. $\text{H}_2\text{SO}_4$ , $\text{Cl}^-$ , $\text{Br}^-$ , $\text{I}^-$ , $\text{NO}_3^-$ , $\text{AgNO}_3$ , $\text{Pb}(\text{CH}_3\text{COO})_2$
4.	Experiment no. 3 Group 3 acidic radical	$\text{BaCl}_2$ , $\text{AgNO}_3$ , $\text{SO}_4^{2-}$ , $\text{B}_4\text{O}_7^{2-}$ , $\text{PO}_4^{3-}$
5.	Experiment no. 4 Group 1 basic radical	$\text{Pb}^{2+}$ , dil HCl, KI, $\text{K}_2\text{CrO}_4$
6.	Experiment no. 5 Group 2 basic radical	$\text{Cu}^{2+}$ , $\text{Cd}^{2+}$ , $\text{Bi}^{3+}$ , dil HCl, $\text{H}_2\text{S}$ , $\text{NaOH}$ , $\text{NH}_4\text{OH}$
7.	Experiment no. 6 Group 3 basic radical	$\text{Al}^{3+}$ , $\text{Fe}^{3+}$ , $\text{Fe}^{2+}$ , $\text{Cr}^{3+}$ , $\text{NaOH}$ , $\text{NH}_4\text{OH}$ , $\text{NH}_4\text{Cl}$ .
8.	Experiment no. 7 Group 4 basic radical	$\text{Zn}^{2+}$ , $\text{Mn}^{2+}$ , $\text{Co}^{2+}$ , $\text{Ni}^{2+}$ , $\text{NaOH}$ , $\text{NH}_4\text{OH}$ , $\text{NH}_4\text{Cl}$ , $\text{H}_2\text{S}$ , $\text{K}_3[\text{Fe}(\text{CN})_6]$
9.	Experiment no. 8 Group 5 basic radical	$\text{Ca}^{2+}$ , $\text{Sr}^{2+}$ , $\text{Ba}^{2+}$ , $\text{NaOH}$ , $\text{NH}_4\text{OH}$ , $\text{NH}_4\text{Cl}$ , $(\text{NH}_4)_2\text{CO}_3$ , $\text{K}_2\text{CrO}_4$ , $\text{K}_2\text{Cr}_2\text{O}_7$
10.	Experiment no. 9 Group 6 basic radical Group 6 basic radical	$\text{Na}^+$ , $\text{K}^+$ , $\text{Mg}^{2+}$ , $\text{NH}_4^+$ , $\text{NaOH}$ , $\text{NH}_4\text{OH}$ , $\text{NH}_4\text{Cl}$ , $(\text{NH}_4)_2\text{CO}_3$ ,

## 1.2 General and Physical Chemistry (201CHEM-4)

Serial	Experiment	Materials (Tools / instruments / Chemicals)	
1.	Group separation of 1st basic radical group	1st basic radical group salts, reagents.	Conical flask , beakers , tubes , filter papers, holders, heater ,vacuum gas chamber
2.	Group separation of 2nd basic radical group	2nd basic radical group salts, reagents.	Conical flask , beakers , tubes , filter papers, holders, heater ,vacuum gas chamber
3.	Group separation of 3rd basic radical group	3rd basic radical group salts, reagents.	Conical flask , beakers , tubes , filter papers, holders, heater ,vacuum gas chamber
4.	Group separation of 4th basic radical group	4th basic radical group salts, reagents.	Conical flask , beakers , tubes , filter papers, holders, heater ,vacuum gas chamber
5.	Group separation of 5th basic radical group	5th basic radical group salts, reagents.	Conical flask , beakers , tubes , filter papers, holders, heater ,vacuum gas chamber
6.	Group separation of 6th basic radical group	6th basic radical group salts, reagents.	Conical flask , beakers , tubes , filter papers, holders, heater ,vacuum gas chamber
7.	Identification Of Inorganic Mixtures	All basic radical group salts, reagents.	Conical flask , beakers , tubes , filter papers, holders, heater, vacuum gas chamber

## 2. Practical Organic Chemistry

### 2.1. Aliphatic Organic Chemistry (231CHEM-3)

Serial	Experiment	Chemicals	Tools / instruments
1.	General Safety	Safety tools in laboratory , Mandatory signs, Safety signs.	
2.	Simple Liquid Organic Compounds: Identification of alcohols	Ethanol, Methanol, Glycerol, Potassium dichromate, Sodium carbonate, Iodine solution, Sodium hydroxide, Borax	Test tube, test tube holder, water bath, Beakers, hot plates.

		solution, Ph. Ph indicator.	
3.	Simple Liquid Organic Compounds: Identification of Aldehydes, and Ketones	Formaldehyde, Acetaldehyde, Schiff reagent, Sodium hydroxide, Iodine solution, 2,4-Dinitrophenylhydrazine, Fehling's solution, Sodium nitroprusside. Test tube, test tube holder, water bath, Beakers, hot plates.	Test tube, test tube holder, water bath, Beakers, hot plates.
4.	Simple Liquid Organic: Identification of Liquid acids and aniline	Acetic acid, Formic acid, Aniline, Sodium carbonate, Potassium permanganate, Sodium nitrite, 2-Naphthol, Sodium hydroxide, Hydrochloric acid, Bromine water.	Test tube, test tube holder, water bath, Beakers, hot plates.
5.	Identification of solid Carboxylic acids	Oxalic acid, tartaric acid, citric acid, benzoic acid, phthalic acid, salicylic acid, Sodium carbonate, Calcium chloride, Ferric chloride, phenol, Sodium hydroxide, Sulphuric acid.	Test tube, test tube holder, water bath, Beakers, hot plates.
6.	Identification of Salts of Carboxylic acids (Metallic and Ammonium salts)	Salts of (Oxalic acid, tartaric acid, citric acid, benzoic acid, phthalic acid, salicylic acid), Sodium carbonate,	Test tube, test tube holder, water bath, hot plates.

		Calcium chloride, Ferric chloride, phenol, Sodium hydroxide, Sulphuric acid.	
7.	Identification of aniline salts and urea	Aniline hydrochloride, Aniline sulphate, Urea, Sodium hydroxide, Sodium nitrite, Hydrochloric acid, Silver nitrate, Barium chloride, Coupper sulphate.	Test tube, test tube holder, water bath, hot plates.
8.	Identifications of carbohydrates	Glucose, fructose, lactose, maltose, sucrose, starch, Sulphuric acid, 1-Naphthol, Barfoied reagent, Fehling` test, Iodine solution,	Test tube, test tube holder, water bath, microscope hot plates.
9.	Scheme of Identification of Unknown liquid organic sample	Any of studied liquid and all the its required reagents , Test tube, test tube holder, water bath, hot plates.	
10.	Scheme of Identification of Unknown Solid organic sample	Any of studied solid organic samples and all the its required reagents, Test tube, test tube holder, water bath, hot plates.	

## 2.2 Aromatic Organic Chemistry (232CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)	
1.	General Safety Rules	Safety tools in laboratory, Mandatory signs, Safety signs.	
2.	Identification of Hydrocarbons	Anthracene, Naphthalene, Picric acid in acetone	Test tube, test tube holder, water bath, hot plates.
3.	Identification of alcohols	Benzyl alcohol, Ethanol, n-Butanol, Lucas reagent, Potassium dichromate	Test tube, test tube holder, water bath, hot plates.
4.	Identification of Phenols	Catechol, Resorcinol, Hydroquinone, 2-Naphthol, 1-Naphthol, Pyrogallol, Gallic acid, Ferric chloride, Bromine water, Sodium nitrite, Sodium hydroxide, Aniline.	Test tube, test tube holder, water bath, hot plates.
5.	Identification of Organic acids and its salts	Acids (Oxalic acid, tartaric acid, citric acid, benzoic acid, phthalic acid, salicylic acid) and its salts (Metallic and Ammonium salts), Sodium carbonate, Calcium chloride, Ferric chloride, phenol, Sodium hydroxide, Sulphuric acid.	Test tube, test tube holder, water bath, hot plates.
6.	Identification of Aromatic Amines	Aniline, Aniline salts, p-Toluidine, 1-Naphthyl amine, Diphenyl amine, Ferric chloride, Bromine water, Sodium nitrite, Sodium hydroxide, 2-Naphthol, Silver nitrate, Barium chloride.	Test tube, test tube holder, water bath, hot plates.

7.	Identification of Amides and Imides	Urea, Acetamide, Benzamide, Succinimide, Phthalimide, Sodium hydroxide, Copper sulphate, Resorcinol, Sulphuric acid.	Test tube, test tube holder, water bath, hot plates.
8.	Identifications of carbohydrates	Glucose, fructose, lactose, maltose, sucrose, starch, Sulphuric acid, 1-Naphthol, Barfoed reagent, Fehling` test, Iodine solution, phenylhydrazine hydrochloride, Sodium acetate.	Test tube, test tube holder, water bath, hot plates.
9.	Tests for physical and Chemical properties: Ignition Test, Heating with soda-lime test, Treatment with 20% NaOH Test and Treatment with Conic H <sub>2</sub> SO <sub>4</sub> Test. Nitration Test, Acidity test, Solubility and reverse precipitation Test, FeCl <sub>3</sub> Test	Benzoic acid, phthalic acid, glucose, starch, resorcinol, 1-naphthol, ammonium oxalate, sodium benzoate, 20% NaOH soln., conc. H <sub>2</sub> SO <sub>4</sub> , soda lime, conc. H <sub>2</sub> SO <sub>4</sub> , conc. HCl, conc. HNO <sub>3</sub> , sod. carbonate, FeCl <sub>3</sub> .	Test tube, test tube holder, water bath, hot plates.
10.	Combination of compounds containing (C, H, O). Identification of Carbohydrates, Carboxylic acids, Phenols, Aldehydes-Ketones, Metallic salts and Hydrocarbon.	Glucose, fructose, lactose, maltose, sucrose, starch, Oxalic acid, tartaric acid, citric acid, benzoic acid, phthalic acid, salicylic acid and its metallic salts, resorcinol, quinol, catechol, 1-naphthol, 2-naphthol, conc. HCl, barfoed reagent, iodine soln., Fehling reagents, FeCl <sub>3</sub> soln., CaCl <sub>2</sub> soln., picric acid soln., 2,4-dinitrophenylhydrazine, Schiff`s reagent	Test tube, test tube holder, water bath, hot plates.

11.	Combination of compounds containing (C, H, O, N). Identification of Ammonium salts of acids, Amide, Imides and Amines.	Amm. (oxalate, tartarate, citrate, benzoate, phthalate, salicylate), urea, phthalimide, FeCl <sub>3</sub> soln., CaCl <sub>2</sub> soln., sod. hydroxide soln., conc.H <sub>2</sub> SO <sub>4</sub> , conc. HCl	Test tube, test tube holder, water bath, hot plates.
12.	Combination of compounds containing (C, H, O, N, S).	Aniline sulphate, sulphanilic acid, thiourea, conc. HCl, sod. nitrite soln., 2-naphthol soln., sod. hydroxide	Test tube, test tube holder, water bath, hot plates.
13.	Combination of compounds containing (C, H, O, N and halogens).	Aniline hydrochloride, conc. HCl, sod. nitrite soln., 2-naphthol soln., sod. hydroxide soln. Test tube, test tube holder, water bath, Bunsen flame	Test tube, test tube holder, water bath, hot plates.

## 2.4 Organic Reaction Mechanisms (335CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)	
1.	Introduction of course 1-Definition of safety in the laboratory 2- Identification of Carboxylic acid organic compounds 3- Identification of carbohydrate. 4-Identification of hydrocarbon. 5- Identification of phenols, 6- Identification of salts of acid (amm. And metallic) 7- Identification of Base	Flame distilled water - test tubes - Beaker- flask	Carboxylic acids, carbohydrates, phenols, Base, HCl, NaOH, Na <sub>2</sub> CO <sub>3</sub>
2.	Acid + Acid mixture	Filter paper Glass	Aliphatic acid (Oxalic acid- Tartic acid- (Citric acid- Succinic acid-) -Aromatic acid

		(Beakers- Separating funnel- Conical flask- Test Tube - Glass plate measuring cylinder.	(Benzoic acid-Phthalic acid- Salicylic acid-Phenyl acetic acid cinnamic acid)  Reagents: (CaCl <sub>2</sub> -FeCl <sub>3</sub> - NH <sub>4</sub> OHNaOH- KMnO <sub>4</sub> - H <sub>2</sub> SO <sub>4</sub> -Con. HCl- NaHCO <sub>3</sub> - Na <sub>2</sub> CO <sub>3</sub> - Resorcinol-Phenol - ethanol-methanol.methanol.
3.	Acid + Phenols	Filter paper Glass (Beakers- Separating funnel- Conical flask- Test Tube - Glass plate measuring cylinder.	Aliphatic acid (Oxalic acid- Tartaric acid- Citric acid- Succinic acid-) -Aromatic acid (Benzoic acid-Phthalic acid- Salicylic acid-Phenyl acetic acid cinnamic acid) -Phenol soluble in water (catechol- Hydro quinone, resorcinol- pyrogallol -Phenols insoluble water (1-naphthol-2- naphthol)  Reagents: (CaCl <sub>2</sub> -FeCl <sub>3</sub> - NH <sub>4</sub> OHNaOH- KMnO <sub>4</sub> - H <sub>2</sub> SO <sub>4</sub> -Con. HCl- NaHCO <sub>3</sub> - Na <sub>2</sub> CO <sub>3</sub> - Resorcinol-Phenol- ethanol-methanol.
4.	Acid + Phenols	Filter paper Glass (Beakers- Separating funnel- Conical flask- Test Tube - Glass plate measuring cylinder.	-Aliphatic acid (Oxalic acid- Tartaric acid- Citric acid- Succinic acid) -Aromatic acid (Benzoic acid-Phthalic acid- Salicylic acid-Phenyl acetic acid cinnamic acid.)  -Phenol soluble in water (catechol-quinolresorcinol- Pyrogallol-Phenols insoluble water (1-naphthol-2- naphthol) - Regents: (CaCl <sub>2</sub> -FeCl <sub>3</sub> - NH <sub>4</sub> OHNaOH- KMnO <sub>4</sub> - H <sub>2</sub> SO <sub>4</sub> -Con. HCl- NaHCO <sub>3</sub> - Na <sub>2</sub> CO <sub>3</sub> - Resorcinol-Phenol-

			ethanol-methanol.
5.	Base + Phenol mixture	Filter paper Glass (Beakers- Separating funnel- Conical flask- Test Tube - Glass plate, measuring cylinder.	Phenol soluble in water (catecholquinol- resorcinol- pyrogallol -Phenols insoluble water (1-naphthol-2- naphthol) – Base (p-Toulidine-1- naphthylamine-2- naphthylamine) - Regents: (CaCl <sub>2</sub> -FeCl <sub>3</sub> -NH <sub>4</sub> OHNaOH- KMnO <sub>4</sub> -H <sub>2</sub> SO <sub>4</sub> -Con. HCl- NaHCO <sub>3</sub> -Na <sub>2</sub> CO <sub>3</sub> - Resorcinol- Phenol- ethanol-methanol.
6.	Acid + Neutral mixture	Filter paper Glass (Beakers- Separating funnel- Conical flask- Test Tube - Glass plate, measuring cylinder.	Aliphatic acid (Oxalic acid- Tartaric acid- Citric acid- Succinic acid-) -Aromatic acid (Benzoic acid-Phthalic acid- Salicylic acid-Phenyl acetic acid cinnamicacid)- Carbohydrate (Glucose- Galatose- Fructose-Lactose- Maltose-Starch- Sucrose)- Hydrocarbons (naphthane- Ancerthane) -Salts of metallic and Salts of ammonium)  -Regent (CaCl <sub>2</sub> -FeCl <sub>3</sub> - NH <sub>4</sub> OH-NaOHKMnO <sub>4</sub> - H <sub>2</sub> SO <sub>4</sub> -Con. HCl-NaHCO <sub>3</sub> - Na <sub>2</sub> CO <sub>3</sub> - Resorcinol-Phenol- 1-naphthol-Feling regents- Berforded Regent- picric acid- acetone- ethanol.
7.	Base + Neutral mixture	Filter paper Glass (Beakers- Separating funnel- Conical flask- Test Tube - Glass plate, measuring	Carbohydrate (Glucose- Galatose-Fructose-  Lactose-Maltose-Starch- Sucrose) - Hydrocarbons(naphthane- Anthracene) -Salts of metallic and Salts of ammonium  -Base (p-Toulidine-1- naphthylamine-2-

		cylinder.	naphthylamine) Regents: (CaCl <sub>2</sub> -FeCl <sub>3</sub> -NH <sub>4</sub> OHNaOH- KMnO <sub>4</sub> -H <sub>2</sub> SO <sub>4</sub> -Con. HCl- NaHCO <sub>3</sub> -Na <sub>2</sub> CO <sub>3</sub> - Resorcinol-Phenol-ethanol-methanol.
8.	Phenol + Neutral mixture	Filter paper Glass (Beakers- Separating funnel- Conical flask- Test Tube - Glass plate, measuring cylinder.	Phenol soluble in water (catechol-quinolresorcinol- Pyrogallol- -Phenols insoluble water (1-naphthol-2- naphthol) - Carbohydrate (Glucose-Galatose- Fructose- Lactose-Maltose-Starch- Sucrose) - Hydrocarbons(naphthane- Ancerthane) -Salts of metallic and Salts of ammonium)-  - Regents: (CaCl <sub>2</sub> -FeCl <sub>3</sub> - NH <sub>4</sub> OHNaOH- KMnO <sub>4</sub> - H <sub>2</sub> SO <sub>4</sub> -Con. HCl- NaHCO <sub>3</sub> - Na <sub>2</sub> CO <sub>3</sub> - Resorcinol-Phenol- - ethanol-methanol.
9.	Neutral + Neural mixture	Filter paper Glass (Beakers- Separating funnel- Conical flask- Test Tube - Glass plate measuring cylinder.	Carbohydrate (Glucose- Galatose-Fructose- Lactose- Maltose-Starch-Sucrose)  -Hydrocarbons (naphthalene- Anthracene) -Salts of metal lic and Salts of ammonium)  Regents: (CaCl <sub>2</sub> -FeCl <sub>3</sub> - NH <sub>4</sub> OHNaOH- KMnO <sub>4</sub> - H <sub>2</sub> SO <sub>4</sub> -Con. HCl- NaHCO <sub>3</sub> - Na <sub>2</sub> CO <sub>3</sub> - Resorcinol-Phenol- ethanol-methanol.

## 2.5 Chemistry of Natural Products (436CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Lab safety	
2.	Preparation of Aspirin	Salicylic acid, acetic anhydride
3.	Preparation of acetanilide	Aniline acetic anhydride
4.	Preparation of Phthalimide	Phthalic anhydride and Urea
5.	Preparation of Phthalyl glycine	Phthalic anhydride and Glycine
6.	Preparation of benzoin	Benzaldehyde and Potassium cyanide
7.	Preparation of benzophenone Oxime	Benzophenone and hydroxylamine hydrochloride
8.	Preparation of 7-hydroxycoumarine	Resorcinol, ethyl acetoacetate, and sulphuric acid
9.	Extraction of caffeine from green tea	Green tea, chloroform, and separating funnel
10.	IR spectra of selected prepared compounds	Infra-Red Spectroscopy apparatus

## 2.6 Applied Organic Chemistry (438CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Phenol formaldehyde resin	Glacial acetic acid, 40% formaldehyde solution, Phenol, conc. HCl. Glass rod, beakers, funnel, heater and filter paper, analytical balance, FTIR.
2.	Urea- Formaldehyde Resins	Urea, Formaldehyde, 35-40 % neutral solution, Oxalic acid, saturated solution. Concentrated ammonia solution Conc. HCl. Flame, Beakers, Test Tubes, Filter papers, Funnels, filtration system, analytical balance
3.	Determination of Saponification Value	Fat, Oil, Fatty acids, Standard N/2 HCl, Alc. KOH and phenolphthalein. Round bottom flask, burette, pipette, water condenser, water bath, analytical balance.
4.	DETERMINATION OF PURITY ANILINE	Aniline hydrochloride, Aniline sulfate,

	SALTS	Standard 0.1 N HCl, and phenolphthalein. burette, pipette, conical flasks and dropper, analytical balance
5.	Determination of the Equivalent Weight of a Carboxylic Acid	<i>Barium hydroxide solution 0.05N, phenolphthalein, carboxylic acids</i> Burette, pipette, conical flasks and dropper, analytical balance
6.	Preparation of para-Red and Dyeing	4-Nitroaniline, 2-naphthol, HCl, Sodium Nitrite, Sodium Hydroxide Beakers, Dropper, Magnetic stirrer, Thermometer, Ice-Bath, Filtration system, Ethanol, Fibers sample, analytical balance, FTIR
7.	Preparation of Soap	Oil, Fat, Sodium hydroxide, Sodium Chloride, Ethanol. Water-bath, thermometer, magnetic stirrer, filtration system, Round-bottomed flask, analytical balance
8.	Synthesis of Biodiesel and studying its properties	Oil, Fat, Potassium hydroxide, Sodium Chloride, Calcium chloride anhydrous, Acetic acid. Water-bath, Separating funnel, Conical flask, analytical balance
9.	Creams	oils, fats, Borax, Mineral oil, water and waxes. Beakers, Water-bath, magnetic stirrer, Thermometer, Filter papers, analytical balance
10.	Preparation of glyptal resin.	phthalic anhydride anhydrous sodium acetate ethylene glycol glycerol analytical balance 2 large test tubes (20- x 150-mm) 1-mL graduated pipette Bunsen burner

		ring stand 2 utility clamps (not rubber coated clamps) FTIR (optional) melting point apparatus (optional) small test tubes or spot plate (optional) assorted solvents such as water, alcohol, acetone,
11	Presentation/Report rubric /Assessment	Theoretical

## 2.7 Principles of Biochemistry (439CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Course Introduction include: -Safety during handling with Chemicals and biological samples. Introduction to the devices used in the laboratory.	Safety tools, and Devices
2.	Carbohydrate detection	Molisch's, Barfoed. Reducing tests, Fehling's, Benedict's, Ammoniacal silver nitrate, Rapid furfural, furfural, Osazone formation and Iodine test
3.	Estimation of the content of reducing sugars using Fehling's and Benedict's test	Fehling's and Benedict's reagent; copper (II) sulfate, potassium sodium tartrate, Potassium hydroxide
4.	Estimation of glucose in serum by phenol-sulphuric acid method	Spectrophotometer, ethanol Phenol, Sulfuric acid, Water bath, Tubes with covers, filter paper, Cones
5.	General tests for proteins	Ninhydrin reagent, copper sulfate in a strong base, sodium hydroxide solution, water bath
6.	Solubility and Precipitation of protein	heavy metals (e.g., $\text{Hg}^{2+}$ , $\text{Pb}^{2+}$ , $\text{Cu}^{2+}$ ), Alkaloidal reagents (e.g., tannate & trichloro acetate), by denaturation (heat coagulation test, strong acids, strong base)
7.	Color reactions of proteins, Biuret test, Millon's test and Reduced sulfur test, Hopkins-Colé test	copper sulfate, sodium hydroxide, Millon's reagent, Hopkins-Colé reagent, $\text{H}_2\text{SO}_4$
8.	Estimation of amino acid	- Ninhydrin

		- titration with potassium hydroxide in the presence of formaldehyde
9.	Properties of fats and oils	Melting point, Crystallization, Viscosity, Density, Solubility, Refractive index, The Saponification number, iodine number, Rancidity
10.	Estimation of triglyceride	4-chlorophenol, Magnesium aspartate, Sodium Azide
	Revision on the theoretical part of the experiments	

### 3. Practical Analytical Chemistry

#### 3.1 Chemistry of Volumetric Analysis (211CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Laboratory safety	None
2.	Solution preparation	Sodium carbonate, sodium chloride, Sulphuric Acid and hydrochloric acid
3.	Determination of normality and strength of unknown sodium hydroxide solution by oxalic acid	Sodium hydroxide, oxalic acid and phenolphthalein
4.	Determination of normality and strength of unknown hydrochloric acid by solution known Sodium hydroxide	Sodium hydroxide, hydrochloric acid, phenolphthalein and methyl orange
5.	Revision	Depending upon the selected experiment
6.	Determination of normality and strength of unknown sodium carbonate solution by standardized Hydrochloric acid solution	sodium carbonate, Hydrochloric acid, phenolphthalein and methyl orange
7.	Determination of normality and strength of unknown potassium permanganate solution by standard oxalic acid solution	potassium permanganate, oxalic acid, Sulphuric Acid

8.	Determination of normality and strength of unknown ammonium ferrous sulphate solution by standard potassium dichromate solution	potassium dichromate solution, ammonium ferrous sulphate, sulphuric acid, phosphoric acid and diphenyl amine
9.	Determination normality and strength of sodium thiosulfate using standard solution of potassium dichromate (iodometric titration)	Sodium thiosulphate, potassium dichromate
10.	Determination normality and strength of magnesium sulphate using standard solution of EDTA (complexometry)	EDTA and magnesium sulphate

### 3.2. Chemistry of Gravimetric Analysis (212CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Laboratory safety, glassware and tools used in gravimetric analysis.	Drying oven, Muffle Furnaces, crucible burette, graduated cylinders, Volumetric flasks, pipettes and conical flasks, water bath, Hot plates, Crucibles, Balances, Water distillation, water deionizer equipments and pH meters
2.	Basic concepts, terminology and gravimetric calculations.	-
3.	Determination of the number of water of crystallization in barium chloride dihydrate	Glassware, Oven, Crucibles, Barium chloride dihydrate and Analytical balance.
4.	Gravimetric analysis of sulphate as $\text{BaSO}_4$	Glassware, Oven, filtration system, sodium sulphate, hydrochloric acid, barium chloride
5.	Gravimetric determination of calcium as calcium Oxide.	Glassware, Furnace, filtration system, hydrochloric acid, calcium salt, ammonia solution, ammonium oxalate.
6.	Gravimetric determination of nickel as nickel dimethylglyoxime	Glassware, Oven, filtration system, hydrochloric acid, nickel chloride, dimethylglyoxime and ammonia solution

7.	Gravimetric determination of lead as lead chromate	Glassware, Oven, filtration system, hydrochloric acid, lead nitrate, potassium chromate, acetic acid, and sodium acetate
8.	Gravimetric determination of iron as ferric oxide	Glassware, Furnace, filtration system, hydrochloric acid, ferrous sulphate, nitric acid, ammonium hydroxide and ammonium nitrate

### 3.3. Chromatographic Analysis (313CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Lab. Experiments' organization and manual design	Data Show.
2.	Safety demonstration	Safety Equipment in the lab.
3.	Separating food colors using paper chromatography	Goggles; gloves and fume hoods. Chromatography paper; Capillary tube to spot samples; Beaker tall-form 500-mL; Watch glass large (to fit beaker); Scissors; Pencil; Ruler. Commercial food colors; Sodium chloride solution, NaCl, 0.1%.
4.	Separating Amino Acids by Thin Layer Chromatography	Goggles; gloves and fume hoods. Thin Layer Chromatography Sheet: (20 x 20 cm) covered with 0.20 mm layer of silica gel 60 (MACHEREY-NAGEL ALUGRAM® Xtra SIL G or similar); Capillary tube to spot samples; Beaker, 1000-mL (Developing Chamber); Watch glass, large (to fit beaker); Spraying bottle for the detecting reagent; Scissors to cut the TLC sheet; Pencil; Ruler. Amino Acids STANDARD solutions of: Lysine, $\beta$ -Alanine, Tryptophan. Developing solvent (Mobile Phase): a mixture of Acetonitrile: water (70:30 vol/vol). Detecting reagent: Ninhydrin solution - 0.3% (w/vol) ninhydrin in n-butyl alcohol containing 3% glacial acetic acid. Drying oven or hair dryer for hot air.

5.	Separation and Identification of Pain-Killing Drugs by Thin Layer Chromatography	<p>Goggles; gloves; face masks and fume hoods.</p> <p>Thin Layer Chromatography Sheet: (20 x 20 cm) covered with 0.20 mm layer of silica gel 60 (MACHEREY-NAGEL ALUGRAM® Xtra SIL G or similar); Capillary tube to spot samples; Beaker, 1000-mL (Developing Chamber); Watch glass, large (to fit beaker); Scissors to cut the TLC sheet; Pencil; Ruler.</p> <p>STANDARD solutions for Active ingredients: Caffeine (6.5 mg/mL); Paracetamol (50 mg/mL); Acetylsalicylic acid (30 mg/mL); Painkiller tablets; Developing solvent (Mobile Phase): a mixture of Ethyl acetate / Hexane / Acetic acid (60:39:1).</p> <p>UV light box with lamp at short wavelength.</p>
6.	Separation of dyes by Column Chromatography	<p>Goggles; gloves; face masks and fume hoods.</p> <p>Chromatography column (400 x 22 mm); Beakers (2), 100-mL; Plastic droppers or Pasteur pipettes; Measuring cylinder, 50-mL; Funnel with wide stem; Pencil (for tapping); Long glass rod to position the cotton wool plug.</p> <p>Dyes Mixture: Mixture of Methyl Orange and Methylene Blue solutions (1:1). Single-compound solutions are prepared in 95% ethanol; Mobile Phase (Elution solvents): FIRST elution solvent: 95% (v/v) Ethanol/Water. SECOND elution solvent: Acetonitrile-Water-Acetic Acid (80:15:5 v/v).</p>
7.	Determination of Caffeine and Benzoic Acid in Soft Drinks by HPLC with UV detector	<p>Goggles; gloves; and fume hoods.</p> <p>HPLC with UV Detector; Ultrasonic bath.</p> <p>Volumetric flasks (2x10 mL); Reagent bottles (1x60 mL); Glass pipette (1x1 mL); Beakers (1x50 mL, 2x25 mL); Syringe Filter (0.2 µm); Plastic syringe (1x2 mL); HPLC glass vial (1x1.5mL).</p> <p>Soft drink sample; Phosphate buffer solution at pH=3 (50 mL); HPLC mobile phase components (Methanol and Phosphate buffer).</p>

8.	Qualitative Separation of Alcohols by Gas Chromatography	Goggles; gloves; and fume hoods. GC with Thermal Conductivity Detector (TCD). Small Vials for the solvents; 10 uL micro syringe. Single-Standard of Alcohols (Methanol, Ethanol, 2-Propanol and 1-Butanol); Mixture of all the four alcohols to examine the separation conditions; Unknown mixture of the above alcohols.
9.	Field trip to a chromatography lab	A bus accommodating the total number of students in addition to 3 instructors.
10.	Experiments review and discussions.	Data show and glassware for demonstration.

### 3.4. Electrochemical Analysis Methods (314CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)	
1.	Safety and regulations	-	-
2.	Potentiometric titration of a strong acid using a strong base	HCl, NaOH and distilled water	Glass burette, beaker 100 mL and pH-meter
3.	Potentiometric titration of a strong acid using a strong base (1 <sup>st</sup> and 2 <sup>nd</sup> derivatization)	-	-
4.	Potentiometric titration of a strong base using a strong acid	HCl, NaOH and distilled water	Glass burette, beaker 100 mL and pH-meter
5.	Potentiometric titration of a strong base using a strong acid base (1 <sup>st</sup> and 2 <sup>nd</sup> derivatization)	-	-
6.	Potentiometric titration of a weak acid using a strong base	Acetic acid, NaOH and distilled water	Glass burette, beaker 100 mL and pH-meter
7.	Potentiometric titration of a weak acid using a strong base (1 <sup>st</sup> and 2 <sup>nd</sup> derivatization)	-	-

8.	Potentiometric titration of a strong base using a weak acid	Acetic acid, NaOH and distilled water	Glass burette, beaker 100 mL and pH-meter
9.	Potentiometric titration of a strong base using a weak acid (1 <sup>st</sup> and 2 <sup>nd</sup> derivatization)	HCl, NaOH and distilled water	Glass burette, beaker 100 mL and pH-meter
10.	Conductometric titration of a mixture of strong acid and weak acid using a strong base	HCl, Acetic acid, NaOH and distilled water	Glass burette, beaker 100 mL and pH-meter

### 3.5. Methods of Instrumental Analysis (415CHEM-4)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Safety measures	-
2.	Introduction to spectroscopic analysis and related calculations	-
3.	Determination of chromium chloride (colored compound) concentration using UV-Vis. Spectrophotometer	Conical flasks, chromium chloride, distilled water UV-Vis. Spectrophotometer
4.	Determination of potassium nitrate (colorless compound) concentration using UV-Vis. Spectrophotometer	Conical flasks, potassium nitrate, distilled water UV-Vis. Spectrophotometer
5.	Spectrophotometric determination of paracetamol in tablets	Conical flasks, paracetamol bulk powder, Panadol tablets, distilled water, methanol, UV-Vis. Spectrophotometer
6.	Spectrophotometric determination of paracetamol in tablets	Conical flasks, paracetamol bulk powder, Panadol tablets, distilled water, methanol, UV-Vis. Spectrophotometer
7.	Spectrofluorimetric determination of eosin yellow dye	Conical flasks, eosin yellow dye, distilled water Spectrofluorimeter
8.	IR- identification of benzoic acid	Benzoic acid, potassium bromide, acetone, IR- spectrometer
9.	Determination of metals concentration using	Nickel standard solutions, water samples containing nickel, nitric acid, distilled water,

	atomic absorption spectrometer	atomic absorption spectrometer
10.	Determination of metals concentration using flame spectrometer	Nickel standard solutions, water samples containing nickel, nitric acid, distilled water , flame spectrometer

## 4. Practical Physical Chemistry

### 4.1 Thermodynamic Chemistry (241CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	<i>Safety and regulations</i>	
2.	<i>The Heat Capacity of the Calorimeter.</i>	<input type="checkbox"/> Styrofoam cups <input type="checkbox"/> Ice <input type="checkbox"/> 100 mL graduated cylinder <input type="checkbox"/> Cardboard lid w/ hole <input type="checkbox"/> DI water <input type="checkbox"/> Burner or hot plate <input type="checkbox"/> Thermometer (-10 to 110 °C) <input type="checkbox"/> 150 mL Beaker <input type="checkbox"/> Watch or Clock <input type="checkbox"/> Thermometer clamp <input type="checkbox"/> 250 mL Beaker <input type="checkbox"/> Centigram balance
3.	<i>Heat of Fusion of ICE.</i>	<input type="checkbox"/> Styrofoam cups <input type="checkbox"/> Ice <input type="checkbox"/> 100 mL graduated cylinder <input type="checkbox"/> Cardboard lid w/ hole <input type="checkbox"/> DI water <input type="checkbox"/> Burner or hot plate <input type="checkbox"/> Thermometer (-10 to 110 °C) <input type="checkbox"/> 150 mL Beaker <input type="checkbox"/> Watch or Clock <input type="checkbox"/> Thermometer clamp <input type="checkbox"/> 250 mL Beaker <input type="checkbox"/> Centigram balance <input type="checkbox"/> metal sample (i.e.: Iron, Copper, Zinc, Aluminum,...)
4.	<i>Specific Heat Capacity of an Unknown Metal.</i>	<input type="checkbox"/> Styrofoam cups <input type="checkbox"/> Ice <input type="checkbox"/> 100 mL graduated cylinder <input type="checkbox"/> Cardboard lid w/ hole <input type="checkbox"/> DI water <input type="checkbox"/> Burner or hot plate <input type="checkbox"/> Thermometer (-10 to 110 °C) <input type="checkbox"/> 150 mL Beaker <input type="checkbox"/> Watch or Clock <input type="checkbox"/> Thermometer clamp <input type="checkbox"/> 250 mL Beaker <input type="checkbox"/> Centigram balance <input type="checkbox"/> metal sample (i.e.: Iron, Copper, Zinc, Aluminum,...)
5.	<i>Heat of Solution of a Salt. (exo- and endo-) thermic dissolution.</i>	<input type="checkbox"/> Styrofoam cup <input type="checkbox"/> Balance <input type="checkbox"/> Thermometer <input type="checkbox"/> 100 mL graduated cylinder <input type="checkbox"/> Anhydrous Sodium acetate, <input type="checkbox"/> Ammonium nitrate, $\text{NH}_4\text{NO}_3$
6.	<i>Heat of Neutralization.</i>	<input type="checkbox"/> Styrofoam cups <input type="checkbox"/> Ice <input type="checkbox"/> 100 mL graduated cylinder

		<input type="checkbox"/> Cardboard lid w/ hole <input type="checkbox"/> DI water <input type="checkbox"/> Burner or hot plate <input type="checkbox"/> Thermometer (-10 to 110 °C) <input type="checkbox"/> 150 mL Beaker <input type="checkbox"/> Watch or Clock <input type="checkbox"/> Thermometer clamp <input type="checkbox"/> 250 mL Beaker <input type="checkbox"/> centigram balance <input type="checkbox"/> NaOH, HCl and CH <sub>3</sub> COOH
7.	<i>Heat of Precipitation.</i>	<input type="checkbox"/> Foam cup <input type="checkbox"/> Thermometer <input type="checkbox"/> Silver nitrate solution <input type="checkbox"/> Sodium chloride solution
8.	<i>Heats of Reaction – Hess’s Law.</i>	<input type="checkbox"/> Styrofoam cup <input type="checkbox"/> Balance <input type="checkbox"/> Thermometer <input type="checkbox"/> 100 mL graduated cylinder <input type="checkbox"/> sodium hydroxide, NaOH <input type="checkbox"/> 1M sodium hydroxide <input type="checkbox"/> 1M Hydrochloric acid <input type="checkbox"/> 0.5M Hydrochloric acid <input type="checkbox"/> Distilled water
9.	<i>The Thermodynamics of Solubility.</i>	<input type="checkbox"/> Solid KNO <sub>3</sub> <input type="checkbox"/> Boiling water bath <input type="checkbox"/> Graduated cylinders <input type="checkbox"/> one 50 mL graduated cylinder with the plastic base removed <input type="checkbox"/> one 25 mL graduated cylinder <input type="checkbox"/> one 10 mL graduated cylinder <input type="checkbox"/> Thermometer or temperature measuring probe <input type="checkbox"/> Large test tube
10.	<i>Spontaneity of Reaction.</i>	<input type="checkbox"/> Solid KNO <sub>3</sub> <input type="checkbox"/> Foam cup <input type="checkbox"/> Graduated cylinders <input type="checkbox"/> Thermometer or temperature <input type="checkbox"/> measuring probe
11.	<i>Determination of Critical Solution Temperature (CST)</i>	<input type="checkbox"/> Test tubes, <input type="checkbox"/> boiling tube as air jacket, <input type="checkbox"/> thermometer (graduated to 0.1°C), <input type="checkbox"/> stirrer, <input type="checkbox"/> beakers, <input type="checkbox"/> phenol, water <input type="checkbox"/> sodium chloride 1N, <input type="checkbox"/> Hot plate.
12.	<i>Phase diagram of 3 Component systems</i>	<input type="checkbox"/> Test tubes, <input type="checkbox"/> thermometer (graduated to 0.1°C), <input type="checkbox"/> stirrer, <input type="checkbox"/> beakers, <input type="checkbox"/> Ethanol / Toluene / Water

## 4.2 Kinetic Chemistry (342CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	<i>Introduction and lab safety</i>	
2.	<i>Determination rate constant</i>	H <sub>2</sub> O <sub>2</sub> , KMnO <sub>4</sub> , sulphuric acid, Potassium

	<i>of hydrogen peroxide decomposition catalyzed by potassium iodide</i>	<i>iodide</i>
3.	<i>Kinetic study of hydrolysis of ethyl acetate catalyzed by acid</i>	<i>Ethyl Acetate, Phenolphthalein, NaOH</i>
4.	<i>Saponification of ethyl acetate.</i>	<i>Ethyl Acetate, Phenolphthalein, NaOH</i>
5.	<i>Determination of rate constant of Iodination of acetone reaction</i>	<i>Acetone, Iodine solution, sulphuric acid, sodium thiosulphate, Starch indicator and Sodium acetate</i>
6.	<i>Effect of temperature on the reaction rate of hydrolysis of ethyl acetate catalyzed by acid and calculation of activation energy</i>	<i>ethyl acetate, Sodium acetate and Hydrochloric acid</i>
7.	<i>Determination of rate constant of persulphate–iodide reaction</i>	<i>Potassium persulphate, Potassium iodide, Sodium thiosulphate and Starch indicator.</i>
8.	<i>Reaction rate of magnesium and hydrochloric acid</i>	<i>Magnesium and Hydrochloric acid</i>

### 4.3 Electrochemistry (344CHEM-3)

Serial	Experiment	4.1 Materials (Tools / instruments / Chemicals)
1.	<i>Safety and regulations</i>	
2.	<i>Determination of cell constant</i>	<i>0.1N KCl, conductivity cell</i>
3.	<i>Determination of equivalent conductance</i>	<i>0.1N KCl, MgSO<sub>4</sub>, monochloric acid, conductivity cell</i>
4.	<i>Activity Series</i>	<i>0.1 M Cu(NO<sub>3</sub>)<sub>2</sub>, 0.1 M Mg(NO<sub>3</sub>)<sub>2</sub>, 0.1 M HCl, 0.1 M Zn(NO<sub>3</sub>)<sub>2</sub>, 0.1 M AgNO<sub>3</sub>, Mg, Cu, Zn</i>
5.	<i>Electrochemical Cells</i>	<i>0.5M Cu(NO<sub>3</sub>)<sub>2</sub>, 0.5M Zn(NO<sub>3</sub>)<sub>2</sub>, 0.5M Pb(NO<sub>3</sub>)<sub>2</sub>, 0.5M KNO<sub>3</sub> rods, DC voltmeter or digital multimeter, copper, zinc, lead.</i>
6.	<i>Galvanic cell creating from environment</i>	<i>Citric acid, Oxalic acid., sheet of copper , sheet of zinc, distilled water, DC voltmeter or digital multimeter, Lemon, Kiwi,.....</i>

7.	<i>Simple galvanic cell using pottery vase or any membrane partition</i>	Zn, Pb, Cu, strips, 0.1M CuSO <sub>4</sub> , 0.1 M Zn(NO <sub>3</sub> ) <sub>2</sub> , 0.1 M Pb(NO <sub>3</sub> ) <sub>2</sub> , 0.1 M FeSO <sub>4</sub> and 0.1 M KNO <sub>3</sub> , DC voltmeter or digital multimeter, porous vase
8.	<i>Investigation of the temperature coefficient Of Galvanic Cell</i>	Copper Sulfate (CuSO <sub>4</sub> ), Zinc Sulfate (ZnSO <sub>4</sub> ), sheet of copper, sheet of zinc, voltmeter or digital multimeter, thermometer
9.	<i>Corrosion</i>	Zn Sheets, NaOH, balance
10.	<i>An Electrolytic Cell: Electrolysis of CuCl<sub>2</sub></i>	0.2 M CuCl <sub>2</sub> . Power supply or 9V batteries
	<i>An Electrolytic Cell: Electroplating</i>	1.0 M CuSO <sub>4</sub> , a copper strip, iron nail, battery or power source,

#### 4.4 Solution Chemistry (344CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Safety and regulations	
2.	Revision on Determination of cell constant	Acids, bases, conductivity cell and conductivity meter
3.	Revision on Determination of equivalent conductance of strong electrolyte	HCl, conductivity cell and conductivity meter
4.	Revision on Determination of equivalent conductance of weak electrolyte	Acetic acid, conductivity cell and conductivity meter
5.	Validation of Debye Huckel theory using Ionic strength Calculations	HCl, NaOH, conductivity cell and conductivity meter
6.	Experimental validation of Kohlrausch's Law for weak electrolytes	HCOOH, Acetic acid, NH <sub>4</sub> OH, conductivity cell and conductivity meter
7.	Experimental validation of Kohlrausch's Law for strong electrolytes	KCl, NaCl, NaOH, conductivity cell and conductivity meter
8.	Experimental Validation	Acetic acid, conductivity cell and conductivity

	of Ostwald's dilution law	meter
9.	Determination of ionization constant of some selected electrolytes	HCl, NH <sub>4</sub> Cl, HCOOH, CH <sub>3</sub> COOH, H <sub>2</sub> SO <sub>4</sub> , conductivity cell and conductivity meter
10.	Safety and regulations	
	Revision on Determination of cell constant	Acids, bases, conductivity cell and conductivity meter

## 5. Practical Inorganic Chemistry

### 5.1 Chemistry of Transition Elements (322CHEM-4)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	<i>Safety and regulations</i>	
2.	Preparation of nickel ammonium sulphate	Glassware - Nickel(II)sulphate hexa-hydrate - Ammonium sulphate.
3.	Determination of nickel as the dimethylglyoximate	Glassware - 1 % alcoholic solution of dimethylglyoxime - HCl (1:1) - Ammonia solution (1:1)
4.	Determination of sulphate as barium sulphate	Glassware - Barium chloride solution (5%) - Concentrated hydrochloric acid.
5.	Calculating of the empirical and the chemical formula of the double salt	
6.	Preparation of copper ammonium sulphate	Glassware - Copper(II)sulphate penta-hydrate. Ammonium sulphate and Acetone.
7.	Determination of copper iodometrically	Glassware - (0.1 N) sodium thiosulphate. Potassium iodide (solid). Starch solution.
8.	Determination of sulphate as barium sulphate	Glassware - Barium chloride solution (5%) - Concentrated hydrochloric acid.
9.	Calculating of the empirical and the chemical formula of the double salt	
10.	Synthesis and characterization of potassium trisoxalatochromate(III) trihydrate	Glassware - Potassium dichromate- Potassium oxalate monohydrate- Oxalic acid dehydrate- Sodium hydroxide- Sulphuric acid- Potassium permanganate- Silver nitrate- Oxalic acid- Ammonium persulphate- Potassium iodide-

		Starch solution-Sodium thiosulphate
--	--	-------------------------------------

## 5.2 Coordination Chemistry (323CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	<i>Safety and regulations</i>	
2.	Introduction about coordination chemistry and the safety in the laboratory.	Periodic table
3.	<u>Direct Titration</u> Determination of Magnesium (II)	**0.01M of EDTA **Buffer ( PH=10)**Soiochrome Black T (E.B.T ) Indicator **Mg <sup>+2</sup> solution
4.	<u>Direct Titration</u> Determination of Zinc (II)	**0.01M of EDTA **Buffer ( PH=10) **Soiochrome Black T (E.B.T ) Indicator **Zn <sup>+2</sup> solution
5.	<u>Direct Titration</u> Determination of cadmium (II)	**0.01M of EDTA **Buffer ( PH=10) **Soiochrome Black T (E.B.T ) Indicator **Cd <sup>+2</sup> solution
6.	<u>Direct Titration</u> Determination of Copper (II)	**0.01M of EDTA **Buffer ( PH=10) **Murexide (Indicator ) **Cu <sup>+2</sup> solution
7.	<u>Direct Titration</u> Determination of Manganese (II)	**0.01M of EDTA **Buffer ( PH=10) **Soiochrome Black T (E.B.T ) Indicator **Mn <sup>+2</sup> solution
8.	<u>Direct Titration Exps.</u> Determination of Lead (II)	**0.01M of EDTA **Buffer ( PH=10) **Soiochrome Black T (E.B.T ) Indicator Pb <sup>+2</sup> solution
9.	<u>Indirect and Back Titration Exps.</u> Determination of Aluminum (III)	**0.01M of EDTA ** 0.01 M Zinc Sulphates **Buffer ( PH=10) **Soiochrome Black T (E.B.T ) Indicator **Al <sup>+3</sup> solution
10.	<u>Indirect and Back Titration Exps.</u> Determination of Nickel (II)	**0.01M of EDTA ** 0.01 M Zinc Sulphates **Buffer ( PH=10) ** Murexide (Indicator ) ** Ni <sup>+2</sup> solution

11.	<i>Substitution Titration Exp.</i> Determination of Calcium	**0.01M of EDTA **Ca <sup>+2</sup> solution **Buffer (PH=10) **Magnesium Complex of EDTA ( Mg-EDTA )
12.	<i>InDirect Titration Exps.</i> Determination of Lead (II)	**0.01M of MgSO <sub>4</sub> **0.01M of EDTA **Buffer ( PH=10) **Soiochrome Black T (E.B.T ) Indicator Pb <sup>+2</sup> solution
13.	Preparation and analysis of monooxalato iron(II) complex	** Glassware. ** Ferrous sulfate ** Ferrous ammonium sulphate . **oxalic acid dihydrate. ** Acetone. ** Ni <sup>+2</sup> solution
14.	Preparation and characterization of potassium trisoxalatochromate(III) trihydrate $K_3[Fe(C_2O_4)] \cdot 3H_2O$	** Glassware. **chromium sulfate ** Potassium dichromate ** Potassium oxalate monohydrate ** Oxalic acid dihydrate ** Sulphuric acid ** Potassium permanganate ** Ammonium persulphate ** H <sub>2</sub> O <sub>2</sub> ** Ethanol
15.	Preparation of potassium cis and trans-diaqua dioxalato chromate (III). Cis & Trans $K[Cr(C_2O_4)_2(H_2O)_2]$	** Glassware. ** chromium sulfate ** Potassium dichromate ** Potassium oxalate monohydrate ** Oxalic acid dihydrate ** Ethanol

### 5.3 Lanthanides & Actinides (424CHEM-3)

Serial	Experiment	Materials (Tools / instruments / Chemicals)
1.	Introduction and lab safety	
2.	The Structure of Atoms	Theoretical calculation
3.	Chemical Periodicity	Theoretical calculation
4.	Chemical Bonding	Theoretical calculation
5.	Exam	Theoretical calculation
6.	Molecular Structure and Covalent Bonding Theories	Theoretical calculation

7.	Coordination Compounds	Theoretical calculation
8.	Revision	

**بعض الأجهزة الطلابية والبحثية المستخدمة في معامل القسم (شطر الطلاب)**

صورة الجهاز	<p>جهاز التحليل الطيفي في مجال الأشعة المرئية وفوق البنفسجية <b>UV-Spectrophotometer</b></p>	اسم الجهاز
<p>يستخدم جهاز الأشعة المرئية وفوق البنفسجية في التحليل الكمي للمواد الكيميائية بصورة رئيسية كما يستخدم أيضا في التحليل الكيفي أو النوعي بدقة عالية.</p>		استخدامات الجهاز
<p>This device used in both quantitative and qualitative analysis in ultra violate and visible spectrum region.</p>		
<p>1. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء ثم انتظر حتي تتم عملية الفحص الذاتي للجهاز. 2. ثبت الطول الموجي المراد القياس عنده. 3. افتح بوابة وضع العينات ثم ضع المحلول الصفري في خليتي العينة 4. قم بتصفير الجهاز (Auto-Zero) 5. قم بوضع العينة المراد قياسها في أحد خلايا وضع العينات ثم قم بقياس الامتصاص للعينة.</p>		طريقة التشغيل
		صورة الجهاز

صورة الجهاز	<p>جهاز صانع الثلج</p> <p><b>Ice maker</b></p>	اسم الجهاز
	<p>يستخدم الجهاز في الحصول علي الثلج اللازم لإجراء وتطبيق بعض الظروف علي التجارب المعملية حيث يعطي الجهاز كمية من الثلج المجروش في زمن قصير نسبيا.</p>	استخدامات الجهاز
	<p>This Device used to produce crushed ice which can be used for some laboratory work.</p>	
	<ol style="list-style-type: none"> <li>1. تاكد من توصيل الجهاز بالماء والكهرباء</li> <li>2. تأكد من أن الماء ينساب بكفاءة عالية.</li> <li>3. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء</li> <li>4. اضغط زر التشغيل (Start)</li> <li>5. سيبدأ الجهاز بتصنيع الثلج في غضون ثلاثون ثانية.</li> <li>6. خذ كمية الثلج التي تكفيك وتأكد من اطفاء الجهاز عند الإنتهاء</li> </ol>	طريقة التشغيل

صورة الجهاز	<p>جهاز تقطير الماء</p> <p><b>Water Distillation Unit</b></p>	اسم الجهاز
	<p>يستخدم الجهاز بصورة رئيسية في إنتاج الماء المقطر اللازم للتجارب المعملية .</p> <p>The device is used to produce distilled water necessary for laboratory work.</p>	استخدامات الجهاز
	<p>1. تأكد من توصيل الجهاز بالماء والكهرباء وأن الماء ينساب بكفاءة.</p> <p>2. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء.</p> <p>3. راقب الجهاز كل فترة زمنية للتأكد من سريان بصورة طبيعية.</p>	طريقة التشغيل

<p>جهاز الامتصاص الذري</p> <p><b>Atomic Absorption spectrophotometer</b></p>	اسم الجهاز
<p>يستخدم الجهاز بصورة رئيسية في التحليل الكمي وتقدير تركيز المعادن في العينات المختلفة خصوصا فلزات المجموعات الانتقالية</p> <p>Atomic absorption spectrophotometer is used for the quantitative determination of samples in free atomic state or in the gaseous state as atomic absorption spectroscopy is based on absorption of light by free metallic ions.</p>	استخدامات الجهاز
<p>1. تأكد من توصيل أنابيب الغاز المشتعل مع الجهاز وأنها تعمل بصورة جيدة بدون تسريب.</p> <p>2. تأكد من مستوي ضغط الغاز في الأنابيب من خلال مقياس ضغط الغاز.</p> <p>3. تأكد من أن مضخة الهواء تعمل بصورة جيدة.</p> <p>4. شغل الجهاز من خلال زر التشغيل ثم قم بتشغيل الكمبيوتر وبرنامج التشغيل للجهاز ثم انتظر حتي تتم عملة الفحص الذاتي.</p> <p>5. قم بتشغيل اللهب بضغط أيقونة (Ignite).</p> <p>6. قم بتشغيل مراوح الشفط في مدخنة الجهاز.</p> <p>7. قم باختيار العنصر المراد قياسه من خلال اختيار اللمية المناسبة له ثم قم بإنشاء نمط التحليل للعينات وللمحاليل القياسية بعد ضبط المحلول الخالي.</p>	طريقة التشغيل

8. تأكد من وضع العينات والمحاليل القياسية بصورة صحيحة ومرتبّة في جهاز الحقن الآلي حسب ما تم وضعه في نمط التحليل.
9. ابدأ عملية القياس للمحاليل القياسية والعينات .
10. بعد اكتمال عملية التحليل تأكد من غسل غرفة الحقن وإخماد اللهب من خلال الضغط علي أيقونة (Extinguish).
11. تأكد من فصل الكهرباء عن الجهاز وإيقاف مروحة الشفط في المدخنة.



صورة  
الجهاز

<p>صورة الجهاز</p>	<p>جهاز انبعاث اللهب <b>Flame Photometer</b></p>	<p>اسم الجهاز</p>
	<p>يستخدم الجهاز بصورة رئيسية في التحليل الكمي وتقدير تركيز المعادن في العينات المختلفة خصوصا فلزات المجموعات الارضية</p> <p>Flame Photometer is used for the quantitative determination of samples by free atoms in the gaseous state.</p>	<p>استخدامات الجهاز</p>
	<ol style="list-style-type: none"> <li>1. تأكد من توصيل انابيب الغاز المشتعل مع الجهاز وأنها تعمل بصورة جيدة بدون تسريب.</li> <li>2. شغل الجهاز من خلال زر التشغيل بعد توصيله بالكهرباء.</li> <li>3. قم بتشغيل الغاز من اسطوانة الغاز ثم قم بحرق اللهب في اعلي المدخنة بواسطة القداحة.</li> <li>4. قم بضبط الفلتر للعنصر المراد قياسه.</li> <li>5. تأكد من وضع العينات والمحاليل القياسية في خرطوم السحب بصورة صحيحة ثم اضغط Enter</li> <li>6. سجل قيمة الانبعاث من الشاشة.</li> </ol>	<p>طريقة التشغيل</p>


<p>صورة الجهاز</p>	<p>فرن التجفيف <b>Drying Oven</b></p>	<p>اسم الجهاز</p>
	<p>يستخدم الجهاز بصورة رئيسية في تجفيف وتسخين العينات عند درجات حرارة لا تتجاوز 300 درجة مئوية.</p>	<p>استخدامات الجهاز</p>
	<p>This device used for drying samples at a temperature less than 300 C° .</p>	<p>طريقة التشغيل</p>
	<p>1. تأكد من توصيل الجهاز بالكهرباء. 2. ضع العينات داخل الجهاز 3. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء 4. اضبط درجة الحرارة المطلوبة من زر الضبط (Mode).</p>	

<p>صورة الجهاز</p>	<p>الميزان الحساس <b>Analytical Balance</b></p>	<p>اسم الجهاز</p>
	<p>يستخدم الجهاز بصورة رئيسية لأخذ أوزان دقيقة حتى 0.0001g للعينات</p>	<p>استخدامات الجهاز</p>
	<p>This device designed to measure small mass in the sub-milligram range.</p>	<p>طريقة التشغيل</p>
	<p>1. تأكد من توصيل الجهاز بالكهرباء. 2. قم بتشغيل الجهاز من زر التشغيل . 3. ضع طبق الوزن علي كفة الجهاز ثم صفر الجهاز بالضغط علي زر (Tare). 4. قم بانزال العينة المراد وزنها بكميات قليلة حتي تصل الي الوزن المطلوب. 5. تأكد من نظافة الجهاز بعد نهاية عملية الوزن</p>	

<p>صورة الجهاز</p>	<p>جهاز حمام الموجات الفوق صوتية</p> <p><b>Ultrasonic Bath</b></p>	<p>اسم الجهاز</p>
	<p>يستخدم الجهاز بصورة رئيسية للمساعدة في إذابة بعض العينات أو طرد الغازات من العينات كما يستخدم في تنظيف بعض الادوات الزجاجية.</p>	<p>استخدامات الجهاز</p>
	<p>It is used to increase the solubility of some samples, degassing and cleaning glassware</p>	
	<ol style="list-style-type: none"> <li>1. تأكد من توصيل الجهاز بالكهرباء</li> <li>2. تأكد من أن كمية الماء في الجهاز في المستوي المطلوب.</li> <li>3. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء.</li> <li>4. اضبط درجة الحرارة المطلوبة من زر الضبط (Mode).</li> <li>5. اضبط درجة الصوت المطلوب من خلال زر ضبط مستوي الصوت.</li> <li>6. اضغط ابدأ (Start)</li> </ol>	<p>طريقة التشغيل</p>

<p>صورة الجهاز</p>	<p>جهاز التحليل الطيفي بالوميض</p> <p><b>Fluorescence spectrometer (Luminescence)</b></p>	<p>اسم الجهاز</p>
	<p>يستخدم جهاز التحليل الطيفي بالوميض في التحليل الكمي للمواد الكيميائية وذلك بتحديد تركيز المواد الكيميائية التي تتميز بخاصية الوميض.</p> <p>The fluormeter used to measure parameters of visible spectrum fluorescence</p>	<p>استخدامات الجهاز</p>
	<ol style="list-style-type: none"> <li>1. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء ثم انتظر حتي تتم عملية الفحص الذاتي للجهاز.</li> <li>2. شغل لمبة الزينون من زر تشغيل اللمبة في الجهاز.</li> <li>3. افتح برنامج تشغيل الجهاز علي الكمبيوتر ثم ثبت الطول الموجي للاثارة والانبعاث للعينة المراد قياسها.</li> <li>4. افتح بوابة وضع العينات ثم ضع المحلول الصفري (Blank) في خلية العينة</li> <li>5. قم بتصفير الجهاز (Auto-Zero)</li> <li>6. قم بوضع العينة المراد قياسها في خلية العينة ثم قم بقياس الانبعاث للعينة بضغط أيقونة Read.</li> </ol>	<p>طريقة التشغيل</p>

<p>صورة الجهاز</p>	<p>جهاز قياس التوصيلية الكهربائية <b>Conductivity meter</b></p>	<p>اسم الجهاز</p>
	<p>قياس التوصيلية الكهربائية هو أداة لقياس التوصيلية الكهربائية في المحاليل. يستخدم هذا الجهاز عادة لمعرفة كمية الأملاح في المحاليل ولها استخدامات عملية في مراقبة كمية الأملاح والشوائب في المصادر المائية</p> <p>Electrical conductivity meter is an instrument for measuring electrical conductivity in solutions. This device is usually used to know the amount of salts in solutions and has practical uses in monitoring the amount of salts and impurities in water sources</p>	<p>استخدامات الجهاز</p>
	<p>(1) يتم غمر خلية التوصيل في prepared standard (2) حدد قائمة MODE باستخدام لوحة المفاتيح. (3) اضغط على مفتاح CAL / CLR. ستتم معايرة الوحدة وفقاً لأقرب معيار. (4) بعد المعايرة يتم قياس العينات بغمر الخلية في العينات والسماح للقراءة بالاستقرار وتسجيل النتيجة. (5) يجب شطف الخلية بالماء منزوع الايونات بين كل عينة لتجنب التلوث ، ورجها لإزالة القطرات الداخلية ، و مسح الخارج قبل الغمر في العينة التالية.</p>	<p>طريقة التشغيل</p>

<p>صورة الجهاز</p>	<p>جهاز قياس الأس الهيدروجيني <b>PH meter</b></p>	<p>اسم الجهاز</p>
	<p>جهاز يستخدم لقياس الاس الهيدروجيني (درجة الحموضة او القاعدية لسائل معين)</p>	<p>استخدامات الجهاز</p>
	<p>A device used to measure pH (the acidity or alkalinity of a specific liquid)</p>	
	<p>1. تشغيل الجهاز 2. تنظيف القطب الكهربائي عن طريق اخلاله من محلول الحفظ وتنظيفه بالماء المقطر في وعاء خارجي ثم تشييفه بالمناديل 3. تجهيز المحاليل المنظمة يفضل عند قياس المحاليل الحمضية استخدام buffer solution ذات الاس الهيدروجيني (4) بينما يفضل استخدام المحاليل المنظمة ذات الاس الهيدروجيني المرتفع لقياس المحاليل القاعدية 4. ضع القطب في العينة وابدأ القراءة واتركها لمدة دقيقة حتى الثبات 5. نظف القطب بعد الاستخدام بشطفه بالماء المقطر وتجفيفه .</p>	<p>طريقة التشغيل</p>

<p>صورة الجهاز</p>	<p>جهاز التحليل بالأشعة تحت الحمراء <b>FT-IR Prestige-21</b></p>	<p>اسم الجهاز</p>
	<p>يستخدم التحليل الطيفي بالأشعة تحت الحمراء في مجموعة متنوعة من التطبيقات ، مثل الكيمياء العضوية العامة ، علم البوليمر ، الأدوية والغذاء.</p> <p>Infrared spectroscopy has traditionally been used in a wide variety of applications, such as general organic chemistry, polymer science, pharmaceuticals, and food.</p>	<p>استخدامات الجهاز</p>
	<ol style="list-style-type: none"> <li>1. صل الجهاز بمصدر التيار الكهربى ثم اضغط زر التشغيل بالواجهة الاماميه ثم قم بفتح التطبيق على جهاز الحاسب الالى ثم انتظر حتى يعطى الجهاز اشارات بدء العمل</li> <li>2. قم بطحن بروميد البوتاسيوم ثم ضع جزءا منه فى القرص القابل ثم ركب القرص الضاغط واكبس العينة بالمكبس لينتج قرص من بروميد البوتاسيوم</li> <li>3. ضع القرص فى الحامل وضعه فى المكان المخصص بالجهاز ثم اضغط القياس بلانك بالتطبيق <b>Back ground</b></li> <li>4. اخلط العينة المراد قياسها مع بروميد البوتاسيوم ثم اطحنها واكبسها كما سبق وضعها بالجهاز ثم اضغط القياس بالتطبيق <b>Sample</b></li> <li>5. قم بطباعة نتيجة التحليل</li> </ol>	<p>طريقة التشغيل</p>

صورة الجهاز	<p>وعاء التسخين</p> <p><b>HEATING MANTLE</b></p>	اسم الجهاز
	<p>ويستخدم هذا الجهاز في معامل الكيمياء العضوية لتسخين القوارير مستديرة القاع لاجراء تجارب تشييد المركبات العضوية</p>	استخدامات الجهاز
	<p>This device is used in organic chemistry laboratories to heat round-bottomed flasks for organic compound synthesis experiments.</p>	استخدامات الجهاز
	<ol style="list-style-type: none"> <li>1. جهز قارورة التجربة ثم ضعها بمكانها في الجهاز ثم ثبتها بالمكثف باحكام</li> <li>2. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء.</li> <li>3. قم بضبط درجة الحرارة المطلوبة لاجراء التجربة 0</li> <li>4. بعد الانتهاء من التجربة قم بايقاف التسخين ثم افصل التيار الكهربائي وانتظر حتى يبرد الجهاز 0</li> <li>5. قم بفك القارورة من المكثف بحذر ومن ثم اخرجها من الجهاز 0</li> </ol>	طريقة التشغيل

<p>صورة الجهاز</p>	<p><b>سخان كهربى بمقلب</b> <b>HOT PLATE</b> <b>WITH STIRRER</b></p>	<p>اسم الجهاز</p>
	<p>يستخدم هذا الجهاز فى معامل الكيمياء العضوية لتسخين التجارب التى تحتاج لدرجة حرارة مرتفعة وأيضا السوائل التى لا يمكن تعريضها للهب المباشر لقابليتها للاشتعال</p>	<p>استخدامات الجهاز</p>
	<p>This device is used in organic chemistry laboratories to heat experiments that require a high temperature, as well as liquids that cannot be exposed to direct flame due to their flammability.</p>	<p>استخدامات الجهاز</p>
	<ol style="list-style-type: none"> <li>1. جهز قارورة التجربة ثم ضعها على سطح الجهاز</li> <li>2. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء.</li> <li>3. قم بضبط درجة الحرارة المطلوبة وضبط سرعة التقلب لاجراء التجربة 0</li> <li>4. بعد الانتهاء من التجربة قم بإيقاف التسخين والتقلب ثم افصل التيار الكهربائى وانتظر حتى يبرد الجهاز 0</li> <li>5. قم بانزال القارورة عن سطح الجهاز بحذر 0</li> </ol>	<p>طريقة التشغيل</p>

صورة الجهاز	<p style="text-align: center;"><b>المجهر الضوئي</b> <b>Microscope</b></p>	اسم الجهاز
	<p>ويستخدم هذا الجهاز في معامل الكيمياء العضوية لتمكين رؤية بعض الاشكال البلورية للمواد الكربوهيدراتية</p>	استخدامات الجهاز
	<p>This device is used in organic chemistry laboratories to enable the visualization of some crystalline forms of carbohydrates</p>	استخدامات الجهاز
	<ol style="list-style-type: none"> <li>1. صل الجهاز بمصدر الكهرباء ثم اضغط زر التشغيل</li> <li>2. قم بوضع المادة البلورية على شريحة نظيفة وجافة ثم ضع الشريحة في مكانها المخصص</li> <li>3. اختر العدسة المناسبة</li> <li>4. قم بضبط قرب العينة من العدسة للحصول على افضل صورة</li> </ol>	طريقة التشغيل

صورة الجهاز	<p>جهاز المبخر الدوراني</p> <p><b>Rotary evaporator</b></p>	اسم الجهاز
	<p>ويستخدم هذا الجهاز في معامل الكيمياء العضوية لازالة المذيبات من العينات بكفاءة بعد عملية الاستخلاص</p>	استخدامات الجهاز
	<p>We can use This device in organic laboratories to evaporate solvents from samples after extraction process</p>	
	<ol style="list-style-type: none"> <li>1. املء الحمام المائي بالماء حتى المستوى المطلوب</li> <li>2. ثبت قارورة العينة في الحمام المائي بفوهة المكثف باحكام وقارورة استقبال المذيب بمكان الاستقبال باحكام</li> <li>3. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء.</li> <li>4. قم بضبط درجة الحرارة المطلوبة لتبخير المذيب وكذلك سرعة الدوران</li> <li>5. انتظر حتى تمام عملية التبخير ثم اغلق التسخين والدوران من الازرار الامامية ثم قم بفك القوارير بحذر</li> </ol>	طريقة التشغيل

صورة الجهاز	حمام مائى <b>Water Bath</b>	اسم الجهاز
	<p>يستخدم الحمام المائى لتسخين المحاليل التى تحتاج لدرجة حرارة أقل والتجارب من درجة غليان الماء كما انه مدعم بمقلب</p> <p>The water bath is used to heat solutions and experiments that require a temperature lower than the boiling point of water, and it is supported by a stirrer.</p>	استخدامات الجهاز
	<ol style="list-style-type: none"> <li>1. يملأ الحمام المائى بالماء حتى المستوى المطلوب</li> <li>2. ثبت قارورة العينة او الدورق فى الحمام المائى</li> <li>3. شغل الجهاز من زر التشغيل بعد توصيله بالكهرباء.</li> <li>4. قم بضبط درجة الحرارة المطلوبة وسرعة الاهتزاز للتقليب عند الحاجة</li> <li>5. انتظر حتى انتهاء التجربة ثم اغلق التسخين والاهتزاز من الازرار الامامية ثم قم بفك القوارير بحذر</li> </ol>	طريقة التشغيل

**بعض الأجهزة المعملية المستخدمة في معامل قسم الكيمياء (شطر الطالبات)**

صورة الجهاز	جهاز قياس درجة انصهار المواد الصلبة <b>(Melting Point apparatus)</b>	اسم الجهاز
	يستخدم هذا الجهاز عادة في معامل الكيمياء في حساب مدى انصهار مادة صلبة معينة.	استخدامات الجهاز
	This device is usually, used in chemistry labs to measure the melting point of a certain solid.	
	1. خذ كمية من المادة الصلبة بواسطة الأنبوبة الشعرية وتكون كمية صغيرة ويتم إنزال المادة الصلبة إلى قاع الأنبوبة الشعرية. 2. تثبت الأنبوبة الشعرية في المكان المخصص لها. 3. خذ قراءة الجهاز عدة مرات فإذا كان الفرق بين درجات الحرارة المأخوذة (2-1) درجة مئوية أي أن المادة الصلبة نقية وخالية من الشوائب وإذا كان الفرق (3-4) درجات مئوية فهذا يعني أن المادة غير نقية وبها شوائب	طريقة التشغيل

<p>صورة الجهاز</p>	<p>جهاز الحمام المائي بهزاز <b>Water Bath Shaker</b></p>	<p>اسم الجهاز</p>
	<p>يستخدم هذا الجهاز في معامل الكيمياء لتسخين العينة حسب درجة الحرارة المطلوبة.</p> <p>This product is used in chemistry laboratories to heat samples according to the required temperature</p>	<p>استخدامات الجهاز</p>
	<p>- تأكد أن المنطقة المحيطة نظيفة وجافة - قم بتوصيل مصدر الطاقة مباشرة ولا تقم بتوصيل قابسين أو أكثر بمأخذ واحد. - تأكد أن الماء في المستوي المطلوب وعال بما يكفي لتغطية عنصر التسخين أثناء التشغيل. - قم بتبديل حمام الماء بعد ذلك. - اضبط أدوات التحكم في درجة الحرارة على درجة الحرارة المطلوبة وانتظر حتى يظهر منظم الحرارة أنه قد تم تسخينه بدرجة كافية قبل البدء. - عند التسخين يرجى ادخال العينات الخاصة بك بعناية، سيحافظ مستشعر درجة الحرارة على حرارة الماء. بعد الاستخدام يرجى إزالة العينات وإيقاف الحمام المائي. - افصل التيار الكهربائي عن الجهاز، ولا تقم مطلقاً بفصل سلك الطاقة عند فصل القابس حيث يؤدي هذا إلى دخول حمل إضافي إلى النظام ويؤدي إلى تعطله.</p>	<p>طريقة التشغيل</p>

صورة الجهاز	فرن التجفيف <b>Dry oven</b>	اسم الجهاز
	<p>يستخدم هذا الجهاز عادة في معامل الكيمياء لتجفيف العينة حسب درجة الحرارة المطلوبة.</p>	
	<p>This device is usually used in chemistry laboratories to dry the sample according to the required temperature.</p>	<p>استخدامات الجهاز</p>
	<ul style="list-style-type: none"> <li>- قم بتوصيل الجهاز بمصدر الطاقة مباشرة.</li> <li>- قم بتشغيل الجهاز من مفتاح التشغيل.</li> <li>- قم بضبط درجة الحرارة حسب الدرجة المطلوبة.</li> <li>- ضع العينة المراد تجفيفها بحرص باستخدام الملقط الخاص بالفرن.</li> <li>- تأكد من تجفيف العينة جيداً.</li> <li>- قم بإغلاق الجهاز من زر التشغيل بعد انتهاء الوقت المحدد للتجفيف.</li> </ul>	<p>طريقة التشغيل</p>

صورة الجهاز	الميزان الكهربائي <b>Electric Balance</b>	اسم الجهاز
	<p>يستخدم هذا الجهاز عادة في وزن العينات.</p>	<p>استخدامات الجهاز</p>
	<p>This device is usually used for weighing samples.</p>	
	<p>- يجب وضع الميزان في غرفة منفصلة وباردة نظيفة، وأن يوضع الميزان على طاولة ثابتة مستوية.</p> <p>- التأكد من أن الميزان بحد ذاته متوازن.</p> <p>- تشغيل الجهاز من مفتاح التشغيل</p> <p>- يصفر الجهاز بعد تشغيله وذلك بالضغط على مفتاح التصفير.</p> <p>- قم بعملية الوزن وأنت واقف أمام الميزان مباشرة.</p> <p>- أمسك الأوزان والمواد المراد وزنها بملقط وضعها في كمنتصف كفة الميزان.</p> <p>- لا توزن الأجسام وهي حارة.</p> <p>- لا تحمل الميزان وزنا أعلى من طاقته.</p> <p>- لا تضع المواد الكيميائية أو الرطوبة فوق كفة الميزان مباشرة.</p> <p>- دون قيمة الوزن بكل دقة لا تحاول أن تقرب قيمة الأرقام.</p> <p>- عند الانتهاء من عملية الوزن يتم تنظيف الميزان بفرشاة خاصة لإزالة أية بقايا من المادة التي تم وزنها والقيام بتغطية الميزان بغطاء بلاستيكي منعا لوصول الأتربة إليه.</p>	<p>طريقة التشغيل</p>

صورة الجهاز	<p>السخان بمقلب مغناطيسي</p> <p><b>Hot plate &amp; stirrer</b></p>	اسم الجهاز
	<p>يستخدم عادة لتسخين العينة عند درجة الحرارة المطلوبة، كما يستخدم المقلب المغناطيسي في تحريك و خلط العينة بفعل مغناطيس.</p>	استخدامات الجهاز
	<p>It is usually used to heat the sample at the required temperature, and the magnetic stirrer is used to stir and mix the sample with the action of a magnet.</p>	
	<ul style="list-style-type: none"> <li>- تأكد من وصول التيار الكهربائي الى السخان.</li> <li>- يتم تشغيل السخان من مفتاح التشغيل.</li> <li>- يتم ضبط درجة الحرارة حسب درجة الحرارة المطلوبة.</li> <li>- بعد الانتهاء من استخدام السخان قومي بإغلاقه من مفتاح التشغيل وفصله عن التيار الكهربائي.</li> <li>- يتم استخدام المقلب المغناطيسي بنفس طريقة التشغيل مع إضافة مغناطيس للعينة للتحريك.</li> </ul>	طريقة التشغيل

صورة الجهاز	جهاز قياس درجة الحموضة (PH meter)	اسم الجهاز
	<p>يستخدم هذا الجهاز عادة في معامل الكيمياء في قياس درجة الحموضة في المحاليل.</p>	استخدامات الجهاز
	<p>This device is usually, used in labs to measurement chemistry the PH of solutions.</p>	
	<p>- تشغيل الجهاز من مفتاح التشغيل - قومي بغسل الإلكترود بالماء المقطر ثم جففيه تماماً. -توضع كمية من محلول معلوم الـ PH في كأس نظيف جاف سعته 50 مل و عادة تستخدم محاليل منظمة. - وضع الكترود الجهاز في المحلول (تجنب التماس الالكترود مع قاعدة الكأس). - ضع الالكترود في المحلول ثم حرك المحلول بتأني تجنباً لتحطيم الالكترود ثم لاحظ قراءة الجهاز. -عدل قراءة الجهاز بواسطة المنظم الخاص لتصبح متساوية للقيمة الحقيقية للمحلول المنظم. -كرر الخطوات 4،5 بعد قطع التيار الكهربائي وإيصاله مرة أخرى مع تعديل قراءة الجهاز في كل مرة لحين استقرار القراءة. - افصل التيار الكهربائي عن الجهاز ثم ارفع المحلول واغسل جيداً بالماء المقطر لعدة مرات ثم جففه تماماً بورق التنشيف</p>	طريقة التشغيل

صورة الجهاز	<p>جهاز التوصيلية الكهربائية</p> <p><b>CONDUCTIVITY METER</b></p>	اسم الجهاز
	<p>يستخدم هذا الجهاز عادة في معامل الكيمياء في قياس التوصيلية الكهربائية في المحاليل.</p> <p>This device is usually, used in chemistry labs to measurement of electrical conductivity in solutions.</p>	استخدامات الجهاز
	<ul style="list-style-type: none"> <li>- قم بتشغيل الجهاز</li> <li>- ثم قم بغسل الالكترود بالماء المقطر ثم جفف تماما</li> <li>- قم بمعايرة الجهاز باستخدام المحاليل المنظمة معلومة الأس الهيدروجيني.</li> <li>- ضع الكترود في المحلول المراد قياس الأس الهيدروجيني له (تأكد من ان الالكترود في وسط المحلول لا يلامس قاع الكأس).</li> <li>سجل القراءة ثم أعد نفس الخطوات لجميع محاليلك غير المعلومة.</li> <li>- قم بغسل الالكترود بالماء المقطر ثم جففه تماما.</li> <li>- قم بإغلاق الجهاز.</li> </ul>	طريقة التشغيل

صورة الجهاز	جهاز الطرد المركزي (CENTRIFUGE)	اسم الجهاز
يستخدم هذا الجهاز عادة في معامل الكيمياء في فصل العينات.		استخدامات الجهاز
This device is usually, used in chemistry labs to separation of samples.		
<p>- يجب وضع جهاز الطرد المركزي على أرضية صلبة أو منصة صلبة مستوية والسعي للحفاظ على الجهاز في وضع أفقي لتجنب اهتزاز الجهاز أثناء الطرد المركزي.</p> <p>- قم بالضغط على مفتاح التشغيل ، ثم قم بتثبيت الجهاز على الإعدادات المطلوبة ، ثم ضع العينة مع ميزان توازن الدرج على حامل العينة الدوارة وإذا لم تتواجد متوازنة العينة يجب استبدال الفرع بماء متساوي الكتلة ؛ يجب وضع أنبوب الطرد المركزي بشكل متناظر في الغلاف لمنع الجسم من الاهتزاز. أغلق الغطاء.</p> <p>- اضغط على زر اختيار الوظيفة لضبط المتطلبات المختلفة: درجة الحرارة والسرعة والوقت والتسارع والتباطؤ.</p> <p>- عند بدء تشغيل جهاز الطرد المركزي، يجب تغطية الغطاء العلوي للطرد المركزي قبل بدء تشغيله ببطء.</p> <p>- اضغط على زر البداية، وسوف تؤدي أجهزة الطرد المركزي المبردة المعلمات أعلاه للعمل، وإغلاق تلقائياً إلى الوقت المحدد.</p> <p>- بعد توقف جهاز الطرد المركزي تمامًا عن الدوران، افتح الغطاء ، واخرج عينة الطرد المركزي ، وامسح الدوار والجدار الداخلي لتجفيف الماكينة بقطعة قماش ناعمة ونظيفة. بعد موازنة درجة الحرارة في حجرة الطرد المركزي مع درجة حرارة الغرفة، يمكن إغلاق الغطاء.</p> <p>- يجب أن يكون الجزء السفلي من غلاف الطرد المركزي مبطنًا بألواح القطن أو أنبوب الاختبار.</p> <p>- بعد الانتهاء من الفصل، أغلق الطرد المركزي أولاً. بعد توقف الطرد المركزي عن الدوران، يمكن فتح غطاء الطرد المركزي لإخراج العينة، ولا يمكن استخدام أي قوة خارجية لإجباره على التوقف عن الحركة.</p>		طريقة التشغيل

- في حالة اهتزاز جهاز الطرد المركزي الكهربائي أو اهتزاز الجسم ، يجب قطع التيار الكهربائي على الفور للقضاء على الصدم فوراً.



صورة  
الجهاز

صورة الجهاز	جهاز فرن الحرق (Muffel Furnaces)	اسم الجهاز
	يستخدم هذا الجهاز عادة في معامل الكيمياء في حرق الرواسب.	استخدامات الجهاز
	This device is usually, used in labs to sediment chemistry burning.	
	<ul style="list-style-type: none"><li>- قم بتوصيل الجهاز بمصدر الطاقة مباشرة.</li><li>- ثم ضبط درجة الحرارة حسب الدرجة المطلوبة.</li><li>- ضع العينة المراد حرقها داخل الفرن</li></ul>	طريقة التشغيل

	<p>باستخدام قفازات خاصة.</p> <ul style="list-style-type: none"> <li>- إحكام إغلاق الباب بشكل صحيح.</li> <li>- انتظار العينة حتى يتم حرقها ثم يتم إخراجها بملقط خاص بفرن الحرق.</li> <li>- قم بإغلاق الجهاز من زر التشغيل بعد الانتهاء.</li> </ul>	
--	---	--

اسم الجهاز	جهاز الاسبكتروفوتومتر (spectrophotometer)	صورة الجهاز
استخدامات الجهاز	<p>يستخدم هذا الجهاز عادة في معامل الكيمياء لقياس كمية الضوء للمادة المستعملة عن طريق طول الموجة التي توجه للجهاز.</p> <p>This device is usually, used in chemistry labs to measure the amount of a substance with the wavelength directed to the instrument.</p>	
طريقة التشغيل	<ul style="list-style-type: none"> <li>- قم بتوصيل الجهاز بمصدر الطاقة مباشرة.</li> <li>- ثم ضبط درجة الطول الموجي حسب نوع المادة المراد قياس تركيزها.</li> <li>- القيام بتصفير الجهاز بـ BLANK (محلول قياسي).</li> <li>- وضع العينة المراد قياس الطول الموجي لها وإغلاق باب الجهاز.</li> <li>- التأكد من ثبات القراءات وأخذ القراءة المناسبة.</li> <li>- يتم إغلاق الجهاز وفصله من التيار الكهربائي.</li> </ul>	